

“UTTARPARA-KOTRUNG MUNICIPALITY”.

TendeNotice No. : UKM-Mahamaya Hospital/001/2018-19

Memo No:1/660

Dated: -12-02-2019

The Chairman, Uttarpara- Kotrung Municipality, Dist. Hooghly, West Bengal invites sealed tender from bonafied, reliable and resourceful suppliers for the work supply of machinery for the treatment of patients admitted in Intensive Care Unit at Mahamaya -Sishu-o-Matrimangal Kendra.

Application, in duplicate, enclosing, single photocopy of necessary credentials.

Valid Pan issued by the I.T.Deptt. ,Govt.of India, valid 15 digit goods and Services

Tax payer identification Number(GSTIN)under GST Act,2017 and Trade License etc.

“Tax invoice(s) needs to be issued by the supplier for raising claim under the contract showing separately the tax charged in accordance with the provision of GST Act,2017”As instructed by the Govt. of West Bengal vide Memo NO- 4374 /F(Y) 13.07.2017 Dt. For obtaining permission of issuing tender documents will have to be submitted to the Chairman of this Municipality with due production of original certificates-Credentials/documents relating to successful completion of similar nature of single work within last 3 (three) years valuing at 1/3rd of the amount of tender under any Government/Semi Government Departments. Complete tender documents will be available for the sale of Mahamaya -Sishu-o-Matrimangal Kendra on payment of the cost tender papers as mentioned below.

However the Municipality on preserves the right of regretting any application without showing any cause thereof. Sealed tenders addressed to the Chairman only. The left hand corner of the envelope containing the tender should be marked with the name for which it is tendered.

The authority does not blind itself to accept the lowest tender and reserves the right to accept or reject any or all tenders without assigning any reason whatsoever.

All other terms and conditions may be seen from the detailed tender documents.

Cess will be charged as per Govt. Rule.

All the Tender documents should be filled up properly by the Supplier, failing to which the Tender will considered as Informal.

Earnest Money should be deposited through Bank Draft/Pay Order in favour of

“UTTARPARA-KOTRUNG MUNICIPALITY”.

PAYMENT TERMS: Payment will be made after all statutory deduction as per government rules.GST will be paid separately as applicable and admissible. Payment will be made through e-payment system through ECS/RECS/RTGS/Cheque after execution of due supply as ordered subject to :

On being selected, the successful vendors will have to submit one application to Chairman, *Uttarpara Kotrung Municipality*. Stating the name of the payee/ recipient, Bank account no with MICR code, IFSC of the payee/recipient for making e payment.

No manual payment is allowed to be made as far as practicable

a) Last date of submission of application with all relevant papers

for obtaining permission: - 22/2/2019

b) Purchase of Tender documents - 25/2/2019

c) Last date of dropping the

Tender documents- 27/02/2019 before 1-00 pm

d) Date of opening Tender - 27/02/2019 after 2-00 pm

| Sl no | Name of Work(s) | Earnest money(Rs.) | Cost of T.paper | Time of Completion |
|-------|--|--------------------|-----------------|--------------------|
| 1 | <p>Supply of one Temporary Pacemaker</p> <p><u>Specifications:</u></p> <p>1) Modes: VVI,AAI,VOO,AOO</p> <p>2) Rate: 30-300ppm</p> <p>3) Rapid Atria Pacing: 80-800ppm</p> <p>4) Output amplitude: 0.1mA to 25mA</p> <p>5) Pulse width:1.5ms</p> <p>6) Sensitivity:0.4 to 20mV</p> <p>7) Operating conditions 10C to 40C(50F to 104F)</p> <p>8) Blanking: 200 ms +5/30 ms-after pace & 120ms+2/30ms-after sense</p> <p>8) Dimensions: width 2.63".Height 7.98".Depth 2.00"</p> <p>9) Weight 0.747lbs</p> <p>10) Battery Type : Standard AA batteries(2x)</p> <p>11) Battery Life :Up to 19 days(under typical conditions)*</p> <p>12) Battery Status Indicator: Yes, gauge type to see what power remains</p> <p>13) Low Battery Behaviour: Red LED on gauge indicator</p> <p>14) Operating time during battery change out:30s(minimum)</p> | 2500.00 | 500.00 | 10 Days |

| | | | | |
|---|--|---------|--------|---------|
| | <p>15)Electrode type: Unipolar /bipolar</p> <p>16)Connector compatibility: Hypertronics D01 style connector. Compatible with all Medtronics patient cables for external pacing(compatible also with 3rd party cables that use this connector style)</p> <p>17)Defibrillator Protection: Internal defibrillation discharge up to 50 J(watts seconds)&external defibrillation discharges up to 360 J</p> <p>18)Digital Display: Yes, LCD, with power saving backlight</p> <p>19)Run Away Rate Protection: Yes</p> <p>20)Lock feature to prevent accidental change of parameters: Yes</p> <p>21)Output pacing circuitry: Constant Current Design</p> <p>22)Auto lock Function: Yes after 60seconds of inactivity</p> | | | |
| 2 | <p>Supply of one Defibrillator (Biphasic)</p> <p><u>Specifications:</u></p> <p>1)Current controlled Biphasic Defibrillator with ECG monitoring.</p> <p>2)Energy level 2 to 300 joules with selection steps of 2,3,5,7,10,20,30,50,70,100, 150,200,250,300 joules.</p> <p>3)Should have facility of stores &print pre &post event data up to 24 Events Recording</p> <p>4)LCD Display size should be more than 5"</p> <p>5)Battery backup more than 100 discharge of 300 Joules in single charge</p> <p>6)Charge time less than 10</p> | 3200.00 | 500.00 | 10 Days |

| | | | | |
|--|---|--|--|--|
| | <p>Sec to 300 J with charge ready LED indicator</p> <p>7)Amplitude gain of ECG waveform should be adjustable.</p> <p>8)Should have paddle ECG & Patient cable ECG(1,2,3 Standard leads)</p> <p>9)Display for heart rate with alarm for high or low HR</p> <p>10)Should be supplied</p> <p>2Bottle of Jelly, !2 roll of thermal paper,3 Lead ECG Cable.</p> <p>11)Should Have Inbuilt Noiseless Thermal Printer</p> <p>12)Should be European CE Approved</p> | | | |
|--|---|--|--|--|

CHAIRMAN

UTTARPARA KOTRUNG MUNICIPALITY