

UTTARPARA-KOTRUNG MUNICIPALITY

N.I.Q-15/MSMK/2023-24
Memo No.1/8068

Date:11/01/2024
Date:11/01/2024

Chairman, Uttarpara Kotrung Municipality invites sealed quotations from bonafied agencies/suppliers for purchase of "Monitor(5 para) "to be used in the **new modular OT** of Mahamaya Sishu –O-Matrimangal Kendra. The quotation to be submitted in a sealed envelope mentioning name of work, quotation no. and date at the top of the envelope.

Quotation is to be submitted addressing, Chairman , Uttarpara Kotrung Municipality during office hours as per the scheduled mentioned below:

01. Critical Date:

- (1)Date of uploading of quotation 11/01/2024
- (2)Last date of submission of quotation is 19/01/2024 by 3pm
- (3)Date of opening of quotation is 19/01/2024 at 4pm

02.Criteria to participate in the quotation:-

Agencies having valid GSTIN Number, PAN, Trade License issued by the competent authority, P Tax, having sound credentials, I.T Return (for last quarter of 2021-2022) and Financial Statement for the year 2021-2022(P/L & Balance Sheet) are entitled to participate in the quotation process.

3.Specifications of the Item:

Attached in a seperate sheet

4.Other terms and conditions:-

The quotationer should quote his rate considering the generic name of the product and keeping the conformity with the existing market price of the product.

The quotationer may remain present during opening of quotation. The quotationer should quote his rate in Indian Rupee including all taxes, charring cost, loading-unloading and other cost including installing, commissioning and supervision.

The Small Scale Industries Unit and Registered Co-operative Societies Ltd. Will get preferences as per Tender Policy as mentioned in G.O No -5400F(Y) Dated 25.06.12 and G.O No.-1500F(Y) Dated 25.06.14.

5.Payments only to be made on the basis of availability of fund after supply and proper checking of bills in all respect.

6.For obtaining permission of issuing quotation documents, a prayer will have to be submitted to the Chairman, Municipality with due production of original certificates- Credentials/ documents.

New G. T. Road, Uttarpara, Dist. Hooghly (West Bengal), Pin - 712 258

Visit us at - www.uttarparamunicipality.in

Mail us at - uttarparakotrungmunicipality@gmail.com

Tele : 033 2663 4095 / 7298

UTTARPARA-KOTRUNG MUNICIPALITY

The authority does not bind itself to accept the lowest quotation and reserves the right to accept or reject any or all quotations without assigning any reason whatsoever.



Chairman


Uttarpara Kotrung Municipality

Memo No.1/8068(4)

Date:11/01/2024

Copy forwarded to:

- 1.Finance Officer,UKM
- 2.Executive Officer,UKM
- 3.Notice Board at UKM
- 4.Acting Medical Superintendent(MSMK)



Chairman

Uttarpara Kotrung Municipality

Page-2-

New G. T. Road, Uttarpara, Dist. Hooghly (West Bengal), Pin - 712 258

Visit us at - www.uttarparamunicipality.in

Mail us at - uttarparakotrungmunicipality@gmail.com

Tele : 033 2663 4095 / 7298

MAHAMAYA SISHU-O-MATRIMONGAL KENDRA

Under The Management of - **UTTARPARA-KOTRUNG MUNICIPALITY**
3B, G. T. ROAD (SAKHER BAZAR), BHADRAKALI, HOOGHLY

Email : uttarparakotrungmunicipality@gmail.com • visit us : www.uttarparamunicipality.in

Regional Diagnostic Laboratory CMDA, IPP-VIII • Ultra Sonography Unit / X-Ray Unit

Specification for Multi Parameter Patient Monitor:

Features

- 12.1 inch color display with 800*600 resolution.
- Three Working Modes: Monitoring, Surgery and Diagnosis.
- Simple and friendly operating display interface.
- Multi-display modes suitable for different applications: Standard Interface, Large font, ECG Standard Full Display, OXY, Trend Table, BP Trend.
- Built-in rechargeable Lithium battery up to 3 hours of continuous work
- VGA output for external display

Technique Specification:

ECG

- Lead Mode: 5 Leads (1, II, III, AVR, AVL, AVF, V)
- Gain: 2.5mm/mV, 5.0mm/mV, 10mm/mV, 20mm/mV
- Heart Rate: 15-300 BPM (Adult); 15-350 BPM (Neonatal)

Resolution: 1 BPM

Accuracy: $\pm 1\%$

Sensitivity $> 200 \mu V$ (Peak to peak)

• ST measurement range: $-2.0 \sim +2.0 \text{ mV}$

Accuracy: $-0.8 \text{ mV} \sim +0.8 \text{ mV}$; $+0.02 \text{ mV}$ or $\pm 10\%$, which is greater

Other Range: unspecified

• Sweep speed: 12.5 mm/s, 25mm/s, 50mm/s

• Bandwidth:

Diagnostic: 0.05~130 Hz

Monitor: 0.5~40 Hz

Surgery: 1~20 Hz

SPO2:

- Measuring Range: 0 - 100 %

Resolution: 1%

Accuracy: 70% ~ 100% (+2 %)

Pulse Rate: 20-250 BPM

Resolution: 1 BPM

Chairman
 Uttara Kotrung
 Uttara Kotrung
 Municipality

Dr. Subharin Mukherjee
 (Gen. Med)
 Consultant
 Mahamaya Sishu-o-Matrimongal Kendra

- Accuracy: ± 3 BPM

NIBP:

- Method: Oscillometry

Measure mode: Manual, Auto, STAT

Unit: mmHg, kPa

- Measure and alarm range:

- Adult Mode

SYS 40 ~ 280 mmHg

DIA 10 ~ 225 mmHg

MEAN 20 ~ 240 mmHg

- Pediatric Mode

SYS 40 ~ 220 mmHg

DIA 10 ~ 160 mmHg

MEAN 20 ~ 170 mmHg

- Neonatal Mode

SYS 40 ~ 135 mmHg

DIA 10 ~ 100 mmHg

MEAN 20 ~ 110 mmHg

- Resolution: 1mmHg

- Accuracy: ± 5 mmHg

Optimal Parameters:

- Recorder (Printer)

- Central monitoring system

- Dual IBP

- Etco2

- 12-lead ECG;

Nellcor SpO2;

- Sun Tech NIBP.

TEMP:

- Measure and Alarm Range: 0 ~ 50 C

- Resolution: 0.1C

- Accuracy: 10.1C

Standard Parameters:

- ECG, RESP, TEMP, NIBP, SPO2, PR

RESP:

- Method: Impedance between RA-LL or RALA

- Measurement Range:

Adult: 10 ~ 120 BrPM

Neonatal / Pediatric: 7 ~ 150 BrPM

- Resolution: 1 BrPM

- Accuracy: ± 2 BrPM



Dr. Subhadip Mukherjee
M.B.B.S. (Cal), D.N.B. (Gen.Mod)
Acting Medical Superintendent
Mahamaya Sishu-O-Matri Hospital, Khandra



Chairman
Jttarpura-Kotrung
Municipality