

# UTTARPARA-KOTRUNG MUNICIPALITY

N.I.Q. No.UKM/MSMK/ 03/MSMK/2021-22

Dated - 29.06.21

Memo No- 1/2297

Date-29.06.21

Chairperson ,Board of Administrator, Uttarpara Kotrung Municipality invites sealed quotations from bonafied agencies/suppliers for suppling following items as mentioned in Sl. No-02 below , for use at Mahamaya Sishu -O-Matrimongal Kendra, a health unit under Uttarpara Kotrung Municipality.

The quotation in a sealed envelop , mentioning name of the work/supply , quotation no , and date at the top of the envelop ,be submitted addressing to the Chairperson, Board of Administrator, Uttarpara - Kotrung Municipality during office hours on the scheduled date as mentioned below .

**01. Critical Date:**

- (1) Date of publishing of quotation 29.06.21
- (2) Last date of submission of quotation is 08.07.21, (up to 3pm)
- (3)Date of opening of quotation on 09.07.21 at about 3.30 pm

**02. Specification of the item to be procured: -**

Sl. No	Name of the Item	Specification	Remarks
01.	Syringe Pump	Specification is attached herewith as Annexure -I	
02.	E.C.G Machine	Rechargeable lithium Battery, having backup 90 minutes. Memory :150 pieces of ECG waveform. Having simultaneous display of 3/6/12 leads ECG waveform and can remark the lead sign ,sensitivity paper speed & filter state.	
03.	5 Para Monitor	Specification is attached herewith as Annexure -II	

**03. Criteria to participate in the quotation : -**

Agencies having valid GSTIN number, PAN, Trade License , I.T. Return (for last quarter of 2019-20 ), & Financial Statement for the year 2019-20( P/L& Balance Sheet) are entitled to participate in the quotation process.

**New G. T. Road, Uttarpara, Dist. Hooghly (West Bengal), Pin - 712 258**

Visit us at - [www.uttarparamunicipality.in](http://www.uttarparamunicipality.in)

Mail us at - [uttarparakotrungmunicipality@gmail.com](mailto:uttarparakotrungmunicipality@gmail.com)

Tele : 2663 4095 / 7298 / 3863 Telefax : 2663 3863 / 7298

# UTTARPARA-KOTRUNG MUNICIPALITY

-02-

## 04. Other terms and conditions:-

The quotationer should quote his rate inclusive of all taxes , keeping conformity with the existing market price of the product to be acquired .

The quotationer may remain present during opening of quotation .The quotationer should quote his rate in Indian Rupee including all taxes ,caring cost, loading-unloading and other cost including installation ,commissioning and supervision.

The Small Scale Industries Unit and Registered Co-operative Societies Ltd will get **preferences** as per Tender policy as mentioned in G.O No-5400F(Y) Dated 25.06.12 and G.O No-1500F(Y) Dated 25.06.14 .


05. **Installation** be made by the supplier free of cost at the entitled site of the Hospital. During installation or commissioning if it is found anything defective or irreparable the same will be replaced by the supplier.

06. The suppliers have to impart **on- site training** to Doctor/ Technicians /Paramedical staff at the time of installation and anytime during warranty period

07. The successful bidder would have to enter in an **agreement regarding CAMC** (Comprehensive Annual Maintenance contract) for a period of 3 years(2<sup>rd</sup> to 4<sup>th</sup> ) after expiry of two years **warranty**.

08. **Payments** only be made on the basis of availability of fund after successful installation and proper checking of bills in all respect .

The authority does not bind itself to accept the lowest quotation and reserves the right to accept or reject any or all quotations without assigning any reason whatsoever.



Chairperson

Board of Administrator

UTTARPARA KOTRUNG MUNICIPALITY

Memo No- 1/2297(4)

Date-29.06.21

## Copy forwarded to :

1. Executive Officer ,UKM
2. Finance Officer, UKM
3. Notice Board at UKM



Chairperson

Board of Administrator

UTTARPARA KOTRUNG MUNICIPALITY

New G. T. Road, Uttarpara, Dist. Hooghly (West Bengal), Pin - 712 258

Visit us at - [www.uttarparamunicipality.in](http://www.uttarparamunicipality.in)

Mail us at - [uttarparakotrungmunicipality@gmail.com](mailto:uttarparakotrungmunicipality@gmail.com)

Tele : 2663 4095 / 7298 / 3863 Telefax : 2663 3863 / 7298

# MAHAMAYA SISHU-O-MATRIMONGAL KENDRA

( Under The Management of - **UTTARPARA KOTRUNG MUNICIPALITY** )

**3B, G. T. Road, (Sakher Bazar), Bhadrakali, Hooghly**

Email : [uttarparakotrungmunicipality@gmail.com](mailto:uttarparakotrungmunicipality@gmail.com) • visit us : [www.uttarparamunicipality.in](http://www.uttarparamunicipality.in)

**Regional Diagnostic Laboratory CMDA, IPP - VIII • Ultra Sonography Unit / X-Ray Unit**

## SPECIFICATION OF Syringe pump:

Syringe sizes	5ml, 10ml, 20ml, 30ml, 50/60ml
Flow rate accuracy	±2%
Flow rate	0.1-1500 ml/h
VTBI	0-9999 ml
Dose rate units	More than 15 types
Concentration calculate	Automatically
Parameters enlarge	Support
Purge	Support
Titration	Support
Bolus setting	Manual bolus Programmable bolus
KVO rate	0.1-5.0 ml/h

*Dilip Yadav*



**Dr. Subhadip Mukherjee**  
M.B.B.S. (Cal), D.N.B. (Gen.Med)  
Acting Medical Superintendent  
Mahamaya Sishu-O-Matri Mongal Kendra

# MAHAMAYA SISHU-O-MATRIMONGAL KENDRA

( Under The Management of - UTTARPARA KOTRUNG MUNICIPALITY )

3B, G. T. Road, (Sakher Bazar), Bhadrakali, Hooghly

Email : uttarparakotrungmunicipality@gmail.com • visit us : www.uttarparamunicipality.in

Regional Diagnostic Laboratory CMDA, IPP - VIII • Ultra Sonography Unit / X-Ray Unit

## Specification for 5 Para Monitor:

### 1. General:

Display: Color TFT

Size: 12.1"

Resolution: 800\*600 pixels

Battery: Rechargeable Lead-Acid battery 12V. 2.0 Ah. Rechargeable Lithium ion battery

Battery operation: >120 minutes operating time under normal use and full charge (with additional battery > 240 mins)

Dimension: 318\*264\*152 mm

### 2. ECG/Respiration/HR:

Lead Mode: 3/5/12 leads ECG input

Heart Rate Range: Adult 10-300 bpm Paediatric & Neonate 10-350 bpm

Sensitivity: >0.2 mvp-p

Accuracy: +/-1% or +/-1 bpm, whichever is greater

Lead selection: I,II,III (3 lead mode)

Respiration Method: Impedance variation between RA-LL (R-F)

Respiration Rate Range: 0-150 rpm

### 3. NIBP:

Method: Automatic Oscillometry

Unit: mmHg/Kpa

Range of Measurement: Adult- 10-270 mmHg.

Child- 10-235 mmHg

Neonate- 10-135 mmHg

Mean- 20-235 mmHg

Pulse rate range: 40-240 bpm

Cuff Pressure Range: Adult 0-280 mmHg

Interval: 60,90 minutes, 2,4,8 hours

### 4. Digital SpO<sub>2</sub>:

Saturation Range: 0-100%

Accuracy: +2%(70-100%).. 0-69% unspecified

Pulse rate range: 25-250 bpm

### 5. Temperature:

Range: 25°C- 50°C

### 6. Trend:

168 hours of graphical trend: HR, NiBP, SpO<sub>2</sub>, PR, Resp, EtCO<sub>2</sub>, Tem1,Temp2, AA,N<sup>2</sup>O, O<sup>2</sup>IBP1,IBP2, ST

*Subhadip Mukherjee*

*Subhadip Mukherjee*

Dr. Subhadip Mukherjee  
M.B.S. (Cal), D.N.B. (Gen.Med)  
Acting Medical Superintendent  
Mahamaya Sishu-O-Matri Mongal Kendra